

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130180 ✓
Name	AW Anderson	ID#	244e7 ✓
Date	09/15/10 ✓		
Agency	Seldovia Police Dept.	Phone #	234-7640
Instrument Location	Seldovia PD		
Alco S/N		X 301530 ✓	Target Value .078 ✓
High Pressure		500 PSI	
Alco Test Values		.075 ✓	.077 ✓
1st Alco			2nd Alco
Signature	AW Anderson		CS 9/22/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of NOV, 2010.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130180

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

SEPTEMBER 15, 2010

OPERATOR'S NAME:

ANDERSON AW

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIW DAY VERIFICATION

SUBJECT'S FIRST NAME/MI :

N/A

D.L. #: 1234567890

DEPT/AGENCY: SOVI

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.30 in		
ALCO TARGET	.079	12:57
BLANK TEST	.000	12:58
INTERNAL STANDARD	VERIFIED	12:58
ALCO TO 30.30 in	.075	12:59
BLANK TEST	.000	12:59
SUBJECT SAMPLE	.000	13:00
BLANK TEST	.000	13:01
ALCO TO 30.32 in	.077	13:01
BLANK TEST	.000	13:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

SEPTEMBER 15, 2010

TIME 12:55

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	30.32 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOFGIRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|~!@